

Corrigendum

In tender document Tender Enquiry No. 24/Urology/254(II) 2017-Rish(Admn)

Dated: 26-09-2017

As per schedule, Pre- Bid meeting of "Mini PCNL Set, Flexible Cysto-Nephro-Fiberscope, Pediatric Cystoscope, VIU and Resect scope for Department of Urology" was held on 18-09-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No24/Urology 254(II))/244/2017-Rish(Admn) has been made.

Mini PCNL: -

Point srl no. 21(i) at page no. 18: -

For: - An applicator consisting of sheath and rod may be supplied so as to use with haemostatic agents like Floseal and Surgiflow.- 5pcs.

Read as: - An applicator consisting of sheath and rod may be supplied so as to use with haemostatic agents like Floseal and Surgiflow.- 5pcs. (Should be supplied from any reputed make and CE)

Flexible Cysto-Nephro-Fiberscope:-

Point srl no. 4 at page no. 19: -

For: - Deflection of distal tip: upward 210 degree and downward 140 degree at least.

Read as: - Deflection of distal tip: upward 210 degree and downward 120 degree at least.

Point srl no. 9 at page no. 19: -

For: - Working length 37cms or more with distal tip diameter of 15.5 Fr or less.

Read as: - Working length 37cms or more with distal tip diameter of 16.5 Fr or less.

Pediatric Cystoscope, VIU and Resectoscope Set:-

Point srl no. 4 at page no. 19: -

For: - Should be supplied with 3 Fr grasping forceps, double action jaws, flexible, length 28 cm.

Read as:- Should be supplied with 3 Fr grasping forceps, double action jaws, flexible, length 28 cm. (Should be supplied from any reputed make and CE)

Point srl no. 5 at page no. 19: -

For:- Should be supplied with 3 Fr biopsy forceps, double action jaws, flexible, length 28cm.

Read as:- Should be supplied with 3 Fr biopsy forceps, double action jaws, flexible, length 28cm. (Should be supplied from any reputed make and CE)

Point srl no. 8 at page no. 19: -

For:- Should be supplied with compatible injection needle, rigid, 3Fr, package of 6 nos.

Read as:- Should be supplied with compatible injection needle, rigid, 3Fr, package of 6 nos. (Should be supplied from any reputed make and CE)

Note: Last date of submission of bids in respect of aforesaid tender is hereby extended till **23-10-2017** at 03:00 PM and Technical bid will be opened on **24-10-2017** at 03.30 PM.