

Corrigendum
In tender document
Tender Enquiry No. 24/PMR/305/2017-Rish(Admn)

Dated: 03-04-2018

As per schedule, Pre- Bid meeting of "Tender for Prosthetic Modelling with Data Acquisition and Fabrication System for Department of Physical Medicine and Rehabilitation (PMR)" was held on 06-03-2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/PMR/305/2017-Rish(Admn)" has been made.

Point Srl no. 05 at page no. 18: -

For: - Data acquisition system should not require calibration or markers for data capture. System should work for at least 5 hours with portable power source without any direct power supply.

Read as: - Data acquisition system should not require calibration or markers for data capture. System should work for at least 4 hours with portable power source without any direct power supply.

Point Srl no. 08(a) at page no. 10: -

For: - Please attach balance sheet (duly certified by Chartered Accountant) for last three (3) years (Attach copy of annual minimum turnover which should not be less than 07 crores duly certified by the Chartered Accountant).

Read as: - Please attach balance sheet (duly certified by Chartered Accountant) for last three (3) years (Attach copy of annual minimum turnover which should not be less than 03 crores duly certified by the Chartered Accountant).

Note: - EMD is also accepted in form of Bank Guarantee.