

# **Corrigendum**

## **In tender document**

**Tender Enquiry No. 24/OT/328(III)/2018-Rish(Admn)**

Dated: 08-05-2018

As per schedule, Pre- Bid meeting of "Tender for Multipara Monitors for OT" was held on 26-04-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/OT/328(III)/2018-Rish(Admn) has been made.

**Point Srl no. 02 at page no. 18: -**

**For: -** Parameter Monitored ECG, HR, Respiration rate, SpO2, NIBP.

**Read as: -** Parameter Monitored ECG, HR, Respiration rate, SpO2.

**Point Srl no. 03 at page no. 18: -**

**For: -** Display: Color LED, approx at least 10 inch with wide viewing angle, facility for display of at least 5 waveforms.

**Read as: -** Display: Color LED, approx at least 15 inch with wide viewing angle, facility for display of at least 12 waveforms. Operation through both full touch screen and trim knob.

**Point Srl no. 04 at page no. 18: -**

**For: -** Soft touch keys, durable and easy to clean

**Read as: -** Operation through both full touch screen and trim knob durable and easy to clean.

**Point Srl. no 19 at page no. 18: -**

**For: -** Should be European CE and USFDA Product.

**Read as: -** Should be European CE/USFDA/ISO/BIS Product.