

Corrigendum
In tender document
Tender Enquiry No. 24/OT/328(VIII)/2018-Rish(Admn)

Dated: 08-05-2018

As per schedule, Pre- Bid meeting of "Tender for Transport and Resuscitation Trolley for OT" was held on 27-04-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/OT/328(VIII)/2018-Rish(Admn)" has been made.

Following points is being added in specification: -

1. Should be hydraulically operated.
2. Should have steering plus system for easy transport.
3. Should have dual locking, four wheel brakes and steer paddle.
4. Should have ergonomically designed push handle at head ad foot end.
5. Should have weight capacity of 310 Kg or more.
6. Should have integrated break system & dual display scale.
7. Trolley should be USFDA/ISO.

However, the quantity is tentative it may be increased/decreased from time to time as per the supply orders. This tender is being converted into rate contract for 02 Years.