

**Corrigendum**  
**In tender document**  
**Tender Enquiry No. 24/CTVS/335/2018-Rish(Admn)**

Dated: - 22-05-2018

As per schedule, Pre- Bid meeting of "Tender for Flexible Bronchoscope for Department of CTVS." was held on 07-05-2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No. 24/CTVS/335/2018-Rish(Admn)-RISH (ADMN)" has been made.

**Point Srl 05 at page no. 18: -**

**For:** - Outer diameter should be 6.0mm or less.

**Read as:** - Outer diameter should be 6.2 mm or less.

**Point Srl no. 6 at page no. 18: -**

**For:** - Channel diameter should be 2.0 mm or more. Separate channels for Oxygen and suction

**Read as:** - Channel diameter should be 2.8 mm or more. Separate channels for Oxygen and suction

**Point Srl no. 11 at page no. 18: -**

**For:** - Minimum visible distance should be 3mm or less. Should have telescopic eye

**Read as:** - Deleted.

**Point Srl no. 13-page no. 18: -**

**For:** - Should have scope ID function.

**Read as:** - Deleted.

**Point Srl no. 16 A - page no. 18: -**

**For:** - The equipment should be supplied with the following standard accessories:

- LED Light Source (Xenon short arc Ozone free 300 Watt lamp): -
- Xenon / LED light with scope compatibility having lamp life of at least 500 hours. -  
Emergency halogen/LED light for backup.
- 4 spare bulbs

**Read as:** - Compatible LED Light Source (xenon short arc ozone free 300 watt), Emergency Halogen light/ LED light for backup, 2 spare bulbs.

**Point Srl no. 16 B-page no. 18: -**

**For:** - Compact Trolley

**Read as:** - Imported Trolley required from same manufacturer or equipment.

**Point Srl no. 16 (C) page no. 18: -**

**For:** - Essential Accessories.

**Read as:** - All same except C. **TBNA needle -20 Pcs. Instead of 5 pack.**

**Point Srl no. 16 (F)-page no. 18: -**

**For:** - Software for video-recording and CD-storage of bronchoscopic procedure (optional- Quoted separately).

**Read as:** - Processor/ Medical grade software should be also be supplied with this which should have facility of capturing the image and video for report generation.

**Following points are being added: -**

1. 26 inch or more LCD/LED Medical grade monitor to be supplied also with set.
2. Color Laser printer of Latest brand from standard company.