

Corrigendum

**In tender document
Tender Enquiry No. 24/RT/Urology/254(VI) 2017-Rish(Admn)**

Dated: 26-11-2017

As per schedule, Pre- Bid meeting of "**URS, Large and URS Small for Department of Urology**" was held on 08-11-2017 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry No24/Urology 269(II))/2017-Rish(Admn) has been made.

URS-Large: -

Point Srl No. 01(4th) at page no. 18: -

For:- Single /Dual working channel Incorporated/Detachable Instrument Port with sealing system and quick release lock 2 channels.

Read as:- **Single channel for both irrigation and instrument.**

Point Srl No. 04 at page no. 18: -

For:-

Accessories (for each scope)

Handle1

Membrane seal (pack of 10) ...2

Membrane retainer standard1

Membrane retainer with luer.... 1

Read as:-

Instrument port with sealing system and quick release lock one channel -1

Instrument port with sealing system and quick release lock two channel -1

Seal (pkt of 10) for instrument port-1

Luer lock tube connector-1

Luer lock tube connector with stopcock-1

URS Small.

Point Srl No. 02 at page no. 19: -

For:-

Accessories (for each scope)

Handle1

Membrane seal (pack of 10)2

Membrane retainer standard1

Membrane retainer with luer.... 1

Read as:-

Instrument port with sealing system and quick release lock one channel -1

Instrument port with sealing system and quick release lock two channel -1

Seal (pkt of 10) for instrument port-1

Luer lock tube connector-1

Luer lock tube connector with stopcock-1

NEPHROSCOPE WITH PCNL SET

Following point from page no. 20 is being amended: -

For: - Should be of 24 Fr-26 Fr for compatible telescope 18-22 Fr, autoclavable, with >12 Fr working channel with Standard/Hollow obturator.

Read as: - Should be of 24 Fr-26 Fr for compatible telescope 18-22 Fr, autoclavable, with >12 Fr working channel with Standard/Hollow obturator **with angled eyepiece**

For: - 380 mm-400mm biopsy forceps for collecting samples for biopsy.

Read as: -380 mm-400mm biopsy forceps for collecting samples for biopsy, (**CE & reputed make**)

For: - Sickle knife for use with Nephroscope Set.

Read as: - Knife for Nephroscope Set.

Note: - Last date of submission of bids in respect of aforesaid tender is hereby extended till 12-12-2017 at 03:00 PM and Technical bid will be opened on next day at 03.30 PM.