

Corrigendum-03

In Tender document
Tender Enquiry No. 24/Nuclear Med/353/2018-Rish (Admn)

14-08-2018

After consideration by store purchase and Technical evaluation committee following modification (deletions/additions/replacements) for Tender Enquiry 24/Nuclear Med/353/2018-Rish(Admn)" has been made.

As some specification correction is not incorporated in previous corrigendum, hence you are requested to view the corrected point (s) as mentioned below-

Tender Reference	Previous PET/CT Specification Corrigendum (1)	Corrigendum (3) Please Read as
Point Tender Document Sr. No.-37 Down time penalty Previous Corrigendum Page No- 01 Dated: 26-07-2018	Downtime Penalty clause of Point S. No.37 at Page No.6 (of <i>Primary tender document</i>). If breakdown, more than 48 hours, warrantee of the equipment will be enhanced one week for further 24 hours. This will have ceiling up to 4 weeks (96 Hours), after that Institute will get it repaired from other vendors on cost of equipment supplier, which shall be first raised for deposit said repair amount and then in non-responding, will be recovered from security deposit 10% of cost of equipment from L-1 bidder, with the penalty upto 10% of repair amount. if such thing repeated in next time, bidder will be debarred with forfeiting of security deposit.	<ol style="list-style-type: none">1. The equipment installed should have 95% uptime during the total warranty and CMC period.2. Equipment downtime will be calculated, if repair are not undertaken upto 48 hrs of registration of complaint till equipment is made functional.3. If the equipment downtime is more than 5%, suitable action shall be taken against the supplier including imposition of penalty as under -<ol style="list-style-type: none">i. If service call not attended and breakdown is more than 48 hours, warrantee of the equipment shall be enhanced for one week for further 24 hours.ii. Downtime for subsequent day(s) shall also result in enhancement of warranty for one week /per day of equipment downtime & so on or equivalent amount will be recovered from security deposit.iii. In a condition, when L-1 bidder not responded for two consecutive times, third time, security deposit of bidder shall be forfeited.

"All other contents of PET/CT Specifications of bidding documents remained unaltered"