

**Corrigendum**  
**In tender document**  
**Tender Enquiry No. 24/Neuro/330(II)/2018-Rish(Admn)**

Dated: 08-05-2018

As per schedule, Pre- Bid meeting of "Tender for Instruments for Epilepsy Surgery & MIS Surgeries for Department of Neurosurgery" was held on 01-05-2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/Neuro/330(II)/2018-Rish(Admn) has been made.

1. EMD of the tender for planning station for epilepsy is Rs 50000 (fifty thousand rupees only)
2. EMD of the tender for instruments for MIS is Rs 60000 (sixty thousand only)
3. EMD for diathermy machine is Rs 50000 (fifty thousand only)
4. If a bidder wants to quote all three of above EMD is 1,60,000 (one lac sixty thousand only)

**For planning station for epilepsy-** no change in specifications

**For MIS instruments-**

- a. For **serial numbers 8-15** - silicon grip is preferable.
- b. For Sr no 30**

**For:-** Bayoneted Sucker- 8.

**Read as:-** Bayoneted sucker 8 or 9

**C. For serial number 31**

**For: -** Bayoneted Sucker- 10

**Read as:-** Bayoneted sucker 10 or 11

Rest of specifications remain unchanged.

**For surgical diathermy machine**

Bidder must provide their bipolar "non-stick" forceps and bipolar forceps with irrigation facility only **standard forceps without non-stick capability will not considered.**

Quote along with their unit price- complete range, from small working length to extra-long length suitable for Transsphenoidal surgeries, the supplied catalogue should have clearly marked pictures of quoted products with dimensions and product code number.

Suitable quantity and types will be chosen at time of final order from their catalogue.