

**Corrigendum**  
**In tender document**  
**Tender Enquiry No. 24/Biochemistry/351/2018-Rish(Admn)**

Dated: 14-06-2018

As per schedule, Pre- Bid meeting of "Tender for Automatic Nucleic Acid Extractor for Department of Biochemistry" was held on 31-05-2018 at 03.00 PM, in the tender opening room.

After consideration by Technical Specification Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/Biochemistry/351/2018-Rish (Admn) has been made.

**Point Srl no. 06 at page no. 18: -**

**For: -** Elution: Elution should be heated with provision of 30-100 µl or better.

**Read as: -** Elution: Elution should be with provision of 30-100 µl or better.

**Point Srl no. 19 at page no. 18: -**

**For: -** Free training of the faculty and technicians to handle instrument and software efficiently.

**Read as: -** Free training of the faculty and technicians to handle instrument and software should be provided till the satisfaction of the faculty in charge of the Biochemistry department.

**Point Srl no. 25 at page no. 18: -**

**For: -** The vendor shall have in-house or access to equipment facilities similar to that is to be supplied and should be able to arrange for site visits for carrying out testing on standards and samples to verify the technical specifications and performance of the equipment by a committee constituted for this purpose by the All India Institute of Medical Sciences, Rishikesh.

**Read as: -** The vendor shall have in-house or access to equipment facilities similar to that is to be supplied and should be able to arrange for carrying out testing on standards and samples to verify the technical specifications and performance of the equipment.