

Corrigendum
In tender document
Tender Enquiry No. 24/Radiotherapay/430/2018-Rish(Admn)

Dated: 08-10-2018

As per schedule, Pre- Bid meeting of "[Tender for Flat Panel Detector Based 1000 MA Digital Radiography & Fluoroscopy System for Department of Radiology.](#)" was held on 06-09-2018 at 03.00 PM, in the tender opening room.

After consideration by Store Purchase Committee following modification (deletions/additions/replacements) additions for Tender Enquiry 24/Radiotherapy/430/2018-Rish(Admn)" has been made.

CRITERIA FOR WEIGHTAGE DURING TECHNICAL EVALUATION (MERIT POINT SYSTEM):

The following merit point system for weighing evaluation factors/ criteria will be applied for technical proposals. Minimum qualifying score is 70 out of 100.

50% weightage will be given for the price bid and following formula will be applied for determining the value for money (vfm) for successful bidder (under the provision of para 1.6 of Ministry of Finance, Government of India Manual of procurement 2017).

$$B = \frac{C_{low}}{C} X + \frac{T}{T_{high}} (1 - X)$$

where

C = Evaluated Proposal Price

C_{low} = the lowest of all Evaluated Proposal Prices among responsive Proposals

T = the total Technical Score awarded to the Proposal

T_{high} = the Technical Score achieved by the Proposal that was scored best among all responsive Proposals

X = weight for the Price as specified in the PDS

There were representations from various vendors and following amendments along with quantitative evaluation criteria has been proposed.

CRITERIA FOR WEIGHTAGE DURING TECHNICAL EVALUATION (MERIT POINT SYSTEM):

The following merit point system for weighing evaluation factors/ criteria will be applied for technical proposals. Minimum qualifying score is 70 out of 100.

S NO.	PARAMETERS	MAXIMUM MARKS
1.	Number of installations of the said equipment in the government institutions in past five years. (four marks for each for a maximum of five installations) Atleast one should be institute of national importance	20 + 10
2.	Certification of satisfactory performance of installation in last three years from head of institution or designated authority by him/her (minimum two, 10 mark for each)	20

3.	Product certification ISO certification -8 USFDA Certification -6 CE Certification -6	20
4.	No litigation with the Procuring Agency/Govt. Dept.	10
5.	Compliance with technical specifications (a)Fully compliant-10 (b)Minor deviation- Compliant with minor deviation (upto 10% subject to main function is not affected)-10	20
	TOTAL	100

SNO	TENDER DOCUMENT	AMENDMENTS
1.	Should have provisions of electromagnetic locks with collision protection sensors.	deleted
2.	System should have motor driven longitudinal and horizontal table top movements.	System should have motor driven longitudinal and horizontal table top movements or head to toe coverage by imaging chain movement.
3.	<u>Other accessories</u> Point number 4. Lead apron with stand-6 nos Point number 5 Thyroid shield- 6 nos. Point number 6 Lead goggles – 6 nos.	To be read as lead aprons with necessary stands- total 12 nos. Skirt and vest type They should have complete overlap on the front opening as well as the vest as well as the vest overlap of the skirt for added protection. They should be of the latest technology and should be light weight with 1mm Pb equivalent. To be read as: Thyroid shields-12 nos. To be read as: Lead goggles- 12 nos.

Point Srl no. 37 at page no. 06:-

For: - The equipment installed should be up for 95% of the total warranty time. If the equipment is down for more than 5% suitable action shall be taken against the supplier including imposition of penalty as deemed fit.

Read as: - The equipment installed should have 95% uptime during the total warranty and CMC period.

If the equipment downtime is more than 5%, suitable action shall be taken against the supplier including imposition of penalty.

In a condition when L-1 bidder not responding for two consecutive times, third time, security deposit of bidder shall be forfeited.

Payment Term for Imported Goods: -

- Letter of Credit (LC) will be opened on 100% Cost of equipment. However, it will be released 75% on shipment & 25% after successful and satisfied installation report.

Payment Term for Site Modification: -

- Site modification is Turnkey and payment will be released after due measurement by Engineering Department and recommendation of indenting department.

Note: -

- EMD is also acceptable in form of Bank Guarantee.
- Delivery period of imported items should be within 90 days from the date of opening of letter of Credit (LC).
- Delivery period of Domestic Items should be within 30 days from the date of issue of Purchase Order.